



Medaval rating of Blood Pressure Monitors recommended by the German consumer organisation *Stiftung Warentest*

Monitors are listed according to clinical validation.

Devices without evidence of validation include those without either a published paper proving validation or evidence of possible MDR equivalence to a validated device. Medaval will update devices where scientific proof of validation or MDR equivalence is provided.

Warning: Wrist monitor accuracy requires strict adherence to manufacturer instructions. If the monitor is not positioned correctly, the results can be very inaccurate.

Stiftung Warentest Recommendations: Very Good (Sehr gut) Good (Gut) Satisfactory (Befriedigend) Adequate (Ausreichend) Poor (Mangelhaft)

Blood Pressure Monitors Validated in a Specific Population

Omron RS7 Intelli IT (HEM-6232T-E)

Blood Pressure Monitors with a possible MDR Equivalence to a Validated Device

Aponorm Basis Control Plus 4G

Hartmann Veroval Compact BPU-22

Visomat Comfort 20-40 (24046) (v3)

Braun ExactFit 5 (BP-6200)

Omron X3 Comfort (HEM-7155-EO)

Blood Pressure Monitors without Evidence of Scientific Validation

Beurer BC 21

Braun ExactFit 3 (BUA6150WE)

Medisana BW 320 (51074)

Beurer BM 51 easyClip

Braun iCheck 7 (BPW4500)

Sanitas SBC-22

BoSo Medicus X

Hartmann Veroval Compact BPW-22

Sanitas SBM-22

BoSo Medistar Plus

Medisana BU 535 (51176)

Visit www.medaval.ie for more information